



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AXIS METROLOGY, INC.  
48861 West Road  
Wixom, MI 48383  
Jeff Johnson Phone: 248 773 7065

CALIBRATION

Valid To: January 31, 2026

Certificate Number: 1767.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 5</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Coordinate Measuring Machines (CMMs) <sup>3</sup> –  Linear Displacement Accuracy: X, Y, Z Axis	Up to 60.96 m	(1.0 + 1.5L) µm	ANSI/ASME B89.4.1 – 1997  ASME B89.4.10360.2 – 2008 Section 6.0:  Laser interferometer, standard gage blocks, ball bar, sphere
CMM Used to Measure Linear Dimensions <sup>3</sup> –  Length Measurement Errors	Up to 1000 mm	(0.46 + 2L) µm	ASME B89.4.10360.2 – 2008 Section 6.0:  Standard gage blocks

<sup>1</sup> This laboratory offers field commercial calibration service only.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC Uncertainty due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in meters.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

**AXIS METROLOGY, INC.**

*Wixom, MI*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 23<sup>rd</sup> day of January 2024.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1767.01  
Valid to January 31, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*